

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 03 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130184</u> ✓	
Name <u>JOHN HAMILTON</u>		ID# <u>3192</u>	Date <u>2/24/10</u> ✓
A	Agency <u>PETERSBURG POLICE DEPARTMENT</u>		Phone # <u>772-3838</u>
Instrument Location <u>PETERSBURG</u>			
Alco S/N <u>78268 BMB</u> ✓ <u>130184</u> ✓ <u>2/24/10</u>			
B	Target Value <u>.081</u> ✓		High Pressure <u>500 PSI</u>
Alco Test Values		<u>.079</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>			
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz  
Nita J. Bolz

Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2<sup>nd</sup> day of April, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

FEBRUARY 24, 2010

OPERATOR'S NAME:

HAMILTON, ADAM A

ANALYST'S NUMBER: 3192

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

INSTRUMENT FIRST NAME/INI :

TEST #1 :

TEST AGENCY: BSGI

TEST REPORT: A

TEST TYPE: U

TEST TARGET VALUE: .001

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

TEST RESULT:

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

FEBRUARY 24, 2010

TIME 22:55

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 06/20/01

HEATERS

SAMPLE CHAMBER:

BREATH TUBE:

PARAMETER: 29.57 in

FLOW RATE: 0.2V

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND

BIND